




**FORMATO DE DATOS MENSUAL DE LOS EQUIPOS SECUENCIALES (PARTISOL) PM10 PAY ENERO 1 - 15 DE 2011**

Fecha dd/mm/aa	Numero de filtro	W Filt limp (g)	W Filt usado (g)	W test inic (g)	W test final (g)	PTP (g)	Volumen (m3 ref)	Concentrac iòn (Ug/m3 ref) con Total Min	Act Start Time	Set Stop Time	llo	lhi	Bplo	Bphi	ICA
01/12/2010	P0000098	0,15000	0,15060	0,79045	0,79046	0,0005900	24,0480	25	00:00	00:00	0	50	0	54	22,72
04/12/2010	P0000099	0,15001	0,15059	0,79045	0,79046	0,0005700	24,0480	24	00:00	00:00	0	50	0	54	21,95
07/12/2010	P0000100	0,15011	0,15072	0,79045	0,79046	0,0006000	24,0480	25	00:00	00:00	0	50	0	54	23,10
10/12/2010	P0000101	0,15019	0,15082	0,79045	0,79046	0,0006200	24,0480	26	00:00	00:00	0	50	0	54	23,87
13/12/2010	P0000102	0,15010	0,15065	0,79045	0,79046	0,0005400	24,0480	22	00:00	00:00	0	50	0	54	20,79
15/12/2010	P0000103	0,15005	0,15069	0,79045	0,79046	0,0006300	24,0480	26	00:00	00:00	0	50	0	54	24,26

**Concentración**



**ICA**

